

Kalibrationsblatt**Baumer sensopress****Calibration Sheet**

Strain Transmitter
DSRT 1000-03-250
Article Code: 134515

SERIE-NUMMER

2.07.527179.1


SERIAL-NUMBER

| | | |
|----------------------------|---------------------------------------|----------------------------|
| Kalibrierdatum Sensor | 16.01.2007 | Calibration Date of Sensor |
| Kalibriert durch | M.Manz | Calibrated by |
| Empfindlichkeit | 250 $\mu\epsilon$ = 10V \pm 2% | Sensitivity |
| Messbrücke | 2 x $\frac{1}{4}$ Bridge 350 Ω | Measuring bridge |
| Linearität | \leq 0.4% FS | Linearity |
| Hysterese | \leq 0.4% FS | Hysteresis |
| Temperatur bei Kalibration | 22°C | Calibration Temperature |

Baumer sensopress AG

Hummelstrasse 17
Postfach
8501 Frauenfeld

Visum / Stempel


.....

Signature / Stamp